



# CERTIFICATE

**TÜV Rheinland Immissionsschutz  
und Energiesysteme GmbH**

---

**Manufacturer:** Oopsis AB

**Measuring System:** Oxygen Monitor O2000

**Components:** O<sub>2</sub>

**Test Report:** 936/808017/A 08.12.1999

---

The measurement system fulfils  
the requirements of

QAL 1

according to EN 14181 and EN ISO 14956.

Köln, 2009-05-05

Dr. rer. nat. Peter Wilbring

Dipl.-Ing. Carsten Röllig

**Calculation of overall uncertainty for QAL1 in EN 14181 and EN 15267-3**
**Manufacturer data**

Manufacturer  
 Name of measuring system  
 Serial Number  
 Measuring Principle

Opsis AB  
 Oxygen Monitor O2000  
 98111\_982115 and 98113\_982117  
 in-situ Zirkondioxid-Sonde

**TÜV Data**

Approval Report  
 Date  
 Editor

936/808017/A 08-02-1999  
 2009-05-05  
 Ruth Steinhagen

**Measurement Component**

certificated range  $O_2$  25 Vol.-%

**Evaluation of the cross sensitivity (CS)**

	QE $X_{max,j}$
to 30 Vol.-% Humidity	0.00 Vol.-%
to 300 mg/m³ Carbon monoxide	0.00 Vol.-%
to 15 Vol.-% Carbon dioxide	0.00 Vol.-%
to 50 mg/m³ Methane	0.00 Vol.-%
to 20 mg/m³ Dinitrogen monoxide	0.00 Vol.-%
to 300 mg/m³ Nitrogen monoxide	0.00 Vol.-%
to 30 mg/m³ Nitrogen dioxide	0.00 Vol.-%
to 20 mg/m³ Ammonia	0.00 Vol.-%
to 1000 mg/m³ Sulphur dioxide	0.00 Vol.-%
to 50 mg/m³ Hydrogen chloride	0.00 Vol.-%
Sum of positive cross sensitivities	0.00 Vol.-%
Sum of negative cross sensitivities	0.00 Vol.-%

**Calculation of the combined standard uncertainty**
**Test Value**

	$\Delta X_{max,j}$	$u^2$
Standard deviation from paired measurements under field conditions * $u_{lof}$	0.10 Vol.-%	0.010
Lack of fit $u_{d,z}$	0.03 Vol.-%	0.000
Zero drift from field test $u_{d,s}$	0.00 Vol.-%	0.000
Span drift from field test $u_t$	0.03 Vol.-%	0.000
Influence of ambient temperature at span $u_p$	0.03 Vol.-%	0.000
Influence of supply voltage $u_f$	0.00 Vol.-%	0.000
Cross sensitivity (interference) ** $u_v$	0.00 Vol.-%	0.000
Uncertainty of reference material at 70% of certification range $u_{rm}$	0.35 Vol.-%	0.041

\* The bigger value of: "Repeatability standard deviation at span" or "Standard deviation from paired measurements under field conditions"

\*\* The absolut value of the sum of negativ cross sensitivity is greater than sum of positiv cross sensitivity

Combined standard uncertainty ( $u_c$ )	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.2
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.44

**Relative total expanded uncertainty**

**Requirement of 2000/76/EC and 2001/80/EC**  
 Requirement of EN 15267-3

<b>U in % of the range 25 Vol.-%</b>	<b>1.8</b>
<b>U in % of the range 25 Vol.-%</b>	<b>10,0</b>
<b>U in % of the range 25 Vol.-%</b>	<b>7,5</b>